



March 1, 2018

Re: 18-25, Addendum #1, Chromatography System

To Whom It May Concern:

1Q) Ionization source can be either ESI or APCI. We can sell with both sources, but it is unclear whether both sources should be quoted or if both sources are available. If only one source is required, then I would need to know which source. Our system can utilize both sources.

1A) The desired source for the quote is ESI, but the instrument needs to have the potential to add APCI if needed in the future.

2Q) Mass Range of 50-1000 Dalton seems to be a typo. The systems I am aware of are mass range of 50-800, 50-1200, or 50-2000.

2A) The correct mass range should be 50-1200 Dalton.

3Q) Flash Chromatography there is a general spec to include cartridges with loading accessories. This is extremely vague as there are a wide range of cartridge media and loading accessories needed depending upon sample volume and type.

3A) The instrument should be able to use different size cartridges and the accessories to use the cartridges. As far as the type of cartridges, the bidder can decide what type of cartridge they would offer.

Addendum: Acknowledge the receipt of each addendum by writing on the bid form the number of the addendum and the date received.

Thank you,

Debra Raines
Bid Specialist